

CERTIFICATE

Certificate No: 11-32663 – Report No: 11836BR-11-32663
Project No: 11-32663


Holder: FysioTape B.V.
Midzomerweg 8
7532 SW ENSCHEDE
Nederland

Product : CureTape®
Identification : 11.0087
Date of test : July 2011


On basis of the results of the tests specified below, we declare that the above specimen fulfils the requirements of the TÜV guidelines

NB: This certificate itself shall not be considered as a type examination certificate.

Ing. A.J. Kettelarij
Author



H. Groothuis
Managing Director



Page 1 of 2

Tests performed:


**Certificate No: 11-32663 – Report No: 11836BR-11-32663
Project No: 11-32663**

Properties and test methods	Requirement	Results
(Onishi & Barr) modified Lowry test	≤ 50 µg/g	Pass

Full details of the results can be found in the following report(s):

Reference	Date of issue
Report: 11836BR-11-32663	december-2011

Ing. A.J. Kettelarij
Author



H. Groothuis
Managing Director